



TECHNICAL DATA ON REVERSE ▲

## BLUE VINYL GLOVES POWDER FREE

CODE: VBF2601/02/03/04

- Examination gloves - regular duty
- Beaded Cuff - providing additional strength and preventing liquid roll off
- Latex Free - to reduce the risk of hypersensitivity and allergic reactions
- Blue Vinyl Gloves provide easy identification in food production
- Ambidextrous
- DOP and EHP free
- Powder free

**1FirstHand**  
HYGIENE DISPENSE SYSTEM

**UNIQUE SINGLE  
HYGIENE DISPENSE  
GLOVE SYSTEM**  
REDUCING THE RISK OF  
CROSS CONTAMINATION





# BLUE VINYL GLOVES POWDER FREE

CODE: VBF2601/02/03/04

## DIMENSIONS

Size Palm Width (mm)	Palm Width (mm)	MGM Length (mm)
S	80	247
M	95	242
L	105	252
XL	113	247

## TEST RESULT

	Inspection Level	AQL
Watertight Test	G1	1.5
Watertight Test	G1	2.5
Minor Defects	G1	4.0
Dimensions	S2	4.0
Physical Properties	S2	4.0

## THICKNESS

Location of Thickness Measurement	Single Wall (mm)
Finger	0.76
Palm	0.51
Cuff	0.51

## PHYSICAL PROPERTIES

	Before Aging
Tensile Strength	8 MPa
Elongation (%)	290 MPa
Force at Break (N)	3.6 W

## ORDERING INFORMATION

Code	Description	Size/Colour	Quantity per inner	Quantity per Carton	Inners per Carton
VBF2601	Blue Vinyl Gloves - Powder Free	Small	100	1000	10
VBF2602	Blue Vinyl Gloves - Powder Free	Medium	100	1000	10
VBF2603	Blue Vinyl Gloves - Powder Free	Large	100	1000	10
VBF2604	Blue Vinyl Gloves - Powder Free	X-Large	100	1000	10

### SYMBOLS



### APPROVED BY



### IN ACCORDANCE WITH

- Medical Device Directive - 93/42/EEC Class 1 (2007/47/EC)
- EN455-1,2,3,4 - Medical Gloves for Single use.
- European Personal Protective Equipment Directive 89/686/EEC - CE Category; CAT3 (0493)
- EN420; Protective Gloves; General Requirements
- EN374-1,2,3-Determination of Resistance to Water Penetration. Leak test passed.
- Chemicals tested: NaOH (40%) Level 6, H2SO4 (96%) Level 1, Heptane Level 0
- ISO 9001: 2000 Certified by SGS (UK) and SIRIM (Malaysia)
- ISO 13485: 2003 Certified by SGS (UK)

PRODUCT INFO ON REVERSE